



Certificate of Compliance

Certificate: 70030133

Master Contract: 217333

Project: 70030133

Date Issued: June 29, 2015

Issued to: Leica Biosystems Nussloch GmbH
Heidelbergerstrasse 17-19
Nussloch, 69226
Germany
Attention: Andre Pristat

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by:

Vicki Gui
Vicki Gui, Certifier

PRODUCTS

CLASS 8721 06 - LABORATORY ELECTRICAL EQUIPMENT

CLASS 8721 86 - ELECTRICAL EQUIPMENT FOR LABORATORY USE (Certified to U.S. Standards)

Cold plate, Model HistoCore Arcadia C, Rated 100 V~ or 110-120 V~, 50/60Hz, 5A Max.

Notes:

1. The above model is cord connected, Equipment Class I, Pollution Degree 2, overvoltage category II
2. Mode of operation: Continuous
3. Environmental Conditions: 20 to 30 °C, 2000 m max, 20 to 80%, non-condensing as specified by manufacturer



Certificate: 70030133

Master Contract: 217333

Project: 70030133

Date Issued: June 29, 2015

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CONDITIONS OF ACCEPTABILITY

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) Equipment is only to be installed by manufacturer trained personnel.
- (4) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.
- (5) Equipment is not to be used with flammable liquids.



Supplement to Certificate of Compliance

Certificate: 70030133

Master Contract: 217333

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70030133	June 29, 2015	Original certification.