



# Certificate of Compliance

**Certificate:** 70030131

**Master Contract:** 217333

**Project:** 70030131

**Date Issued:** June 10, 2015

**Issued to:** Leica Biosystems Nussloch GmbH  
Heidelbergerstrasse 17-19  
Nussloch, 69226  
Germany  
Attention: Andre Pristat

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only*



**Issued by:**

*Vicki Gui*  
Vicki Gui, Certifier

## **PRODUCTS**

CLASS 8721 06 - LABORATORY ELECTRICAL EQUIPMENT

CLASS 8721 86 - ELECTRICAL EQUIPMENT FOR LABORATORY USE (Certified to U.S. Standards)

Embedding station, Model HistoCore Arcadia H, Rated 100-120 V~ or 220-240 V~, 50/60 Hz, 10 A Max, 1000 VA.

## **Notes:**

1. The above model is cord connected, Equipment Class I, Pollution Degree 2, Overvoltage Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: 20 to 30 °C, 2000 m max, 20 to 80%, non-condensing as specified by manufacturer.



**Certificate:** 70030131

**Master Contract:** 217333

**Project:** 70030131

**Date Issued:** June 10, 2015

---

### **APPLICABLE REQUIREMENTS**

#### CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

#### UL Standards:

UL Std. No. 61010-1 (3<sup>rd</sup> Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

### **CONDITIONS OF ACCEPTABILITY**

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) Equipment is only to be installed by manufacturer trained personnel.
- (4) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.



## *Supplement to Certificate of Compliance*

**Certificate:** 70030131

**Master Contract:** 217333

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
70030131	June 10, 2015	Original certification.