

## EU DECLARATION OF CONFORMITY

We, **Leica Biosystems Nussloch GmbH, Heidelberger Str. 17-19, 69226 Nussloch, Germany**

declare on our own responsibility, that the medical device

### **Leica ASP6025**

complies with

- the essential requirements of the Directive 98/79/EC of the European Parliament and of the Council of 27 October 1998 on in vitro diagnostic medical devices (OJ L 331, 7.12.1998, p. 1-37), Annex III

EN 61010-2-101:2017

EN ISO 14971:2012

EN 61326-2-6:2013

Device classification: IVD general/other

- Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (OJ L 174, 1.7.2011, p. 88-110)


EN 50581:2012

Quality Management System: Certified according to EN ISO 13485:2016 and ISO 9001:2015

Manufacturing sites: Leica Instruments (Singapore) Pte Limited, 12 Teban Gardens Crescent, Singapore 608924

Leica Microsystems Ltd. Shanghai, Building 1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade, 201206 Shanghai, People's Republic of China

Nussloch, 04.02.2019



Martin Laszak  
Director Customer Service Core Histology



Robert Gropp  
RA/QA Manager

Rev. A