



EC Declaration of Conformity

We herewith declare, in exclusive responsibility, that the instrument

Leica SM2500 – Sliding Microtome

was developed, designed and manufactured to conform with the

- Directive 2006/95/EC of the European Parliament and of the Council (Low Voltage)
- Directive 2004/108/EC of the European Parliament and of the Council (electromagnetic compatibility)

The following harmonized standards were applied:


- **EN 61010-1: 2001 + Corr.: 2002 + Corr.: 2004**
Safety requirements for electrical equipment for measurement, control and laboratory use -
Part 1: General requirements
- **EN 61326: 2006 + Corr.: 2008**
Electrical equipment for measurement, control and laboratory use -
EMC requirements -
Part 1: General requirements
- **EN 1037:1995+A1:2008**
Safety of machinery - Prevention of unexpected start-up;
- **EN ISO 13850:2008 + Cor.:2009**
Safety of machinery - Emergency stop - Principles for design
- **EN ISO 14121-1:2007**
Safety of machinery - Risk assessment - Part 1: Principles

In addition, the following in-house standards were applied:

- **DIN EN ISO 9001: 2000**
Quality management systems - Requirements

Leica Biosystems Nussloch GmbH
Heidelberger Str. 17-19
69226 Nussloch, Germany
October 30, 2009

14 0383 80011


.....
Arnd Kaldowski
President Biosystems Division