

Kreatech™ FISH probes

Product Information Sheet

KI-10718


PTEN (10q23) / SE 10

100 µl

DANGER



FORMAMIDE




Kreatech Biotechnology B.V.
Vlierweg 20
1032 LG Amsterdam
The Netherlands
www.LeicaBiosystems.com

RUO - Research Use Only

Not for use in diagnostic procedures

PI-KI-10718_D2.1

Published Sept 2015



Not to scale

Kreatech™ PTEN (10q23) / SE 10 FISH probe

Introduction: The **PTEN (10q23)** specific FISH probe is optimized to detect the PTEN gene region at region 10q23. The **Satellite Enumeration (SE) 10** FISH probe is included to facilitate chromosome identification.

Critical region 1 (red): The **PTEN (10q23)** gene region probe is direct-labeled with PlatinumBright™550.
Control region (green): The **SE 10** FISH probe is direct-labeled with PlatinumBright™495.

Reagent: Kreatech probes are direct-labeled DNA probes provided in a ready-to-use format. Apply 10 µl of probe to a sample area of approximately 22 x 22 mm.

Please refer to the Instructions for Use for the entire Kreatech FISH protocol.

Kreatech FISH probes are REPEAT-FREE™ and therefore do not contain Cot-1 DNA. Hybridization efficiency is increased and background, due to unspecific binding, is highly reduced.

Interpretation: The **PTEN (10q23) / SE 10** FISH probe is designed as a dual-color assay to detect deletions at 10q23. Deletions involving the PTEN gene region at 10q23 will show one red signal, while the control at the chromosome 10 centromere region will provide 2 green signals. Two single color red (R) and green (G) signals will identify the normal chromosomes 10 (2R2G).

	Normal Signal Pattern	10q23 Deletion
Expected Signals	2R2G	1R2G

References: Cairns et al, 1997, Cancer Res, 57 ; 4997-5000
 Hermans et al, 2004, Genes Chrom Cancer, 39; 171-184

Warning and precautions: In case of emergencies check SDS sheets for medical advice. SDS sheets may be obtained by either contacting Leica Technical Support or visiting www.LeicaBiosystems.com. DNA probes contain formamide which is a teratogen; do not inhale or allow skin contact. Wear gloves and a lab coat when handling DNA probes. All materials should be disposed of according to your institution's guidelines for hospital waste disposal.

Reagent Storage and Handling: Store at 2-8 °C. Reagents should not be used after the expiration date on the vial label.

TECHNICAL SUPPORT Technical support is available at www.LeicaBiosystems.com/service-support/technical-support/ or toll free at 800-248-0123 or via e-mail: kreatech-support@leicabiosystems.com.

CUSTOMER SERVICE Kreatech probes may be ordered through Leica Customer Service toll free at 800-248-0123 or order via e-mail: purchase.orders@leica-microsystems.com.