

Kreatech™ FISH probes

Product Information Sheet

KI-40102


"N25" (22q11) / 22q13 (SHANK3)

100 µl

DANGER



FORMAMIDE




Kreatech Biotechnology B.V.
Vlietweg 20
1032 LG Amsterdam
The Netherlands
www.LeicaBiosystems.com

RUO - Research Use Only

Not for use in diagnostic procedures

PI-KI-40102_D2.1

Published Sept 2015



Not to scale

KI-40102

Kreatech™ “N25” (22q11) / 22q13 (SHANK3) FISH probe

Introduction: The “N25” (22q11) region probe covers the marker “N25” (D22S75) and adjacent region of CLTD (Chlatriin gene region) and DGCR2 (DiGeorge critical region gene 2). The 22q13 (SHANK3) FISH probe is optimized to detect copy numbers of the SHANK3 gene region at 22q13.

Critical region 1 (red): The “N25” (22q11) specific FISH probe is direct-labeled with PlatinumBright™550.
Critical region 2 (green): The 22q13 (SHANK3) specific FISH probe is direct-labeled with PlatinumBright™495.

Reagent: Kreatech probes are direct-labeled DNA probes provided in a ready-to-use format. Apply 10 µl of probe to a sample area of approximately 22 x 22 mm.

Please refer to the Instructions for Use for the entire Kreatech FISH protocol.

Kreatech FISH probes are REPEAT-FREE™ and therefore do not contain Cot-1 DNA. Hybridization efficiency is increased and background, due to unspecific binding, is highly reduced.

Patterns: The “N25” (22q11) / 22q13 (SHANK3) FISH probe is designed as a dual-color assay to detect deletions at 22q11 and 22q13. Deletions involving the DiGeorge “N25” region will show one red signal and two green signals (1R2G). Deletions involving the 22q13 (SHANK3) region will show one green signal and two red signals (2R1G). Two single color red and green signals will identify the normal chromosomes 22 (2R2G).

	Normal Signal Pattern	Del(22q11) “N25”	Del(22q13) SHANK3
Expected Signals	2R2G	1R2G	2R1G

References: DiGeorge, A. M. 1968, Birth Defects Orig. Art. Ser. IV(1): 116-121
Mattei, M.-G et al, 1994, Genomics 23: 717-718
Scambler, P, 2000, Human Molecular Genetics, 9; 2421-2426
Wilson, et al, 2003, J Med Genet 40; 575-584
Luciani, et al, 2003, J Med Genet 40; 690-696

Warning and precautions: In case of emergencies check SDS sheets for medical advice. SDS sheets may be obtained by either contacting Leica Technical Support or visiting www.LeicaBiosystems.com. DNA probes contain formamide which is a teratogen; do not inhale or allow skin contact. Wear gloves and a lab coat when handling DNA probes. All materials should be disposed of according to your institution’s guidelines for hospital waste disposal.

Reagent Storage and Handling: Store at 2-8 °C. Reagents should not be used after the expiration date on the vial label.

TECHNICAL SUPPORT Technical support is available at www.LeicaBiosystems.com/service-support/technical-support/ or toll free at 800-248-0123 or via e-mail: kreatech-support@leicabiosystems.com.

CUSTOMER SERVICE Kreatech probes may be ordered through Leica Customer Service toll free at 800-248-0123 or order via e-mail: purchase.orders@leica-microsystems.com.