

Kreatech™ FISH probes

Product Information Sheet

KI-40104


TBX1 (22q11) / 22q13 (SHANK3)

100 µl

DANGER



FORMAMIDE




Kreatech Biotechnology B.V.
Vlierweg 20
1032 LG Amsterdam
The Netherlands
www.LeicaBiosystems.com

RUO - Research Use Only

Not for use in diagnostic procedures

PI-KI-40104_D2.1

Published Sept 2015



Not to scale

KI-40104

Kreatech™ TBX1 (22q11) / 22q13 (SHANK3) FISH probe

Introduction: The **TBX1 (22q11)** region probe is optimized to detect the TBX1 gene region at 22q11.
The **22q13 (SHANK3)** FISH probe is optimized to detect the SHANK3 gene region at 22q13.

Critical region 1 (red): The **TBX1 (22q11)** specific FISH probe is direct-labeled with PlatinumBright™550.
Critical region 2 (green): The **22q13 (SHANK3)** specific FISH probe is direct-labeled with PlatinumBright™495.

Reagent: Kreatech probes are direct-labeled DNA probes provided in a ready-to-use format.
Apply 10 µl of probe to a sample area of approximately 22 x 22 mm.

Please refer to the Instructions for Use for the entire Kreatech FISH protocol.

Kreatech FISH probes are REPEAT-FREE™ and therefore do not contain Cot-1 DNA. Hybridization efficiency is increased and background, due to unspecific binding, is highly reduced.

Patterns: The **TBX1 (22q11) / 22q13 (SHANK3)** FISH probe is designed as a dual-color assay to detect deletions at 22q11 and 22q13. Deletions involving the TBX1 will show one red signal and two green signals (1R2G). Deletions involving the 22q13 (SHANK3) region will show one green signal and two red signals (2R1G). Two single color red and green signals will identify the normal chromosomes 22 (2R2G).

	Normal Signal Pattern	Del(22q11) TBX1	Del(22q13) SHANK3
Expected Signals	2R2G	1R2G	2R1G

References: DiGeorge, A, 1968, Birth Defects Orig. Art. Ser. IV(1); 116-121
Lindsay et al, 2001, Nature 410; 97-101
Liao et al, 2004, Human Molecular Genetics 13; 1577-1585
Luciani, et al, 2003, J Med Genet 40; 690-696
Wilson et al, 2003, J Med Genet 40; 575-584

Warning and precautions: In case of emergencies check SDS sheets for medical advice. SDS sheets may be obtained by either contacting Leica Technical Support or visiting www.LeicaBiosystems.com. DNA probes contain formamide which is a teratogen; do not inhale or allow skin contact. Wear gloves and a lab coat when handling DNA probes. All materials should be disposed of according to your institution's guidelines for hospital waste disposal.

Reagent Storage and Handling: Store at 2-8 °C. Reagents should not be used after the expiration date on the vial label.

TECHNICAL SUPPORT Technical support is available at www.LeicaBiosystems.com/service-support/technical-support/ or toll free at 800-248-0123 or via e-mail: kreatech-support@leicabiosystems.com.

CUSTOMER SERVICE Kreatech probes may be ordered through Leica Customer Service toll free at 800-248-0123 or order via e-mail: purchase.orders@leica-microsystems.com.